

720350-1 RANGE 90683		CHECK READING AT 25 ±10°C	
	°F	°C	
mV	±101.1mV		5.054 ±0.012mV
J	-346 to 2192	-210 TO 1200	96.1 ±0.3°C
K	-328 TO 2501	-200 TO 1372	123.2 ±0.4°C
T	-405 TO 752	-243 TO 400	116.4 ±0.4°C

ART 135



Versa-Cal TM **CALIBRATOR**

BIDDLE INSTRUMENTS

MADE IN U.S.A.



USE ONLY IN ACCORDANCE WITH INSTRUCTION MANUAL

S/N 90683 PN 720350-1