

 **BARFIELD**

W.O. No.: WO1196195

Date of Certification:	8/19/2024	Certification Due Date:	8/19/2025
Part Number:	101-00901	Received in Tolerance:	NO
Serial Number:	5972	Returned in Tolerance:	YES
Description:	DIGITAL TURBINE TEMP. T/S	Temperature:	74°
Manufacturer:	BARFIELD INC.	Humidity:	51%

Deviation: N/A

Equipment	Serial Number	Last Date Calibrated	Next Calibration Date
4115EB	184D18118	9/23	9/24
3729EB	4485009	7/24	7/25
4782E	78893	10/23	10/24
2398EB	0002	7/24	7/25

Results: RECONNECTED LISTED PART, AND FIXED 25 (OHM) POTENTIOMETER. ALSO  
CALIBRATED AND TESTED W/I MFG SPECS.

**This certificate applies only to item identified and shall not be reproduced without the permission from Barfield Inc.**

F.D. 01045

Rev.6 Issued Date: 07/18/2018