


BARFIELD INC.  
4101 N.W. 29TH ST.  
Miami FL - 33142  
(800) 321-1039



**CERTIFICATE OF CALIBRATION/  
FUNCTIONAL TEST**

W.O. No.: WO1198827

Customer Name: PIUS PARTS

Date of Certification: 10/10/2024 Certification Due Date: 10/10/2025  
Part Number: 101-00901 Received in Tolerance: NO  
Serial Number: 3380 Returned in Tolerance: YES  
Description: DIGITAL TURBINE TEMP. T/S Temperature: 74°  
Manufacturer: BARFIELD INC. Humidity: 51%

Calibration Procedure Used: 20-101-00901

Deviation: N/A

**STANDARDS USED TO CERTIFY EQUIPMENT**

Equipment	Serial Number	Last Date Calibrated	Next Calibration Date
4115EB	184D18118	09/24	09/25
3729EB	4485009	07/24	07/25
4782E	78893	07/24	07/25
2398EB	0002	07/24	07/25

Comments: THE UNIT FAILED TESTS DUE TO THE FOLLOWING ISSUES: CONTAMINATED BOTH PCB  
BOARDS, CABLE ASSEMBLY AND BULKHEAD ASSEMBLY DAMAGED.

Results: REPLACED ALL PARTS LISTED. CALIBRATED AND TESTED UNIT W/I MFG. SPECS.

This is to certify that the above listed part number meets or exceeds all specifications as stated in the referenced procedure at the points tested (unless otherwise noted). Calibration Standards used in the calibration of this product are traceable to the National Institute of Standards and Technology (NIST). Unless otherwise stated the uncertainty of the measurement standards shall not exceed 25% of the acceptable tolerance. This calibration is in accordance with Barfield Inc. Quality Management System (QPM-7.6) which is ISO 9001:2015/AS9100 D certified and compliant with MIL-STD-45662A, ANSI/NCSL Z540.1-1994 (R2002) and ISO 10012.

This certificate applies only to item identified and shall not be reproduced without the permission from Barfield Inc.

Certified by: Claudia Galvez Title/Stamp: Q.C. INSPECTOR C.G. 99910

Dated: 10/10/2024