

BARFIELD INC.  
4101 N.W. 29TH ST.  
Miami FL - 33142  
(800) 321-1039



### CERTIFICATE OF CALIBRATION/ FUNCTIONAL TEST

W.O. No.: WO1232039

Customer Name: PIUS PARTS

Date of Certification: 8/25/2025 Certification Due Date: 8/25/2026  
Part Number: 101-00901 Received in Tolerance: NO  
Serial Number: 6205 Returned in Tolerance: YES  
Description: DIGITAL TURBINE TEMP. T/S Temperature: 74°  
Manufacturer: BARFIELD INC. Humidity: 51%

Calibration Procedure Used: 20-101-00901

Deviation: N/A

#### STANDARDS USED TO CERTIFY EQUIPMENT

Equipment	Serial Number	Last Date Calibrated	Next Calibration Date
4115EB	184D18118	9/24	9/25
4782E	77358	8/25	8/26
3729EB	4485009	7/25	7/26
2398EA	B8119	7/25	10/25

Comments: UNIT FAILED TEST DUE TO DEFECTIVE 45V BATTERY ASSEMBLY CONNECTOR, FET, AND ALSO LITHIUM BATTERY LOW VOLTAGE.

Results: REPLACED ALL LISTED PARTS. CALIBRATED AND TESTED W/I MFG SPECS.

This is to certify that the above listed part number meets or exceeds all specifications as stated in the referenced procedure at the points tested (unless otherwise noted). Calibration Standards used in the calibration of this product are traceable to the National Institute of Standards and Technology (NIST). Unless otherwise stated the uncertainty of the measurement standards shall not exceed 25% of the acceptable tolerance. This calibration is in accordance with Barfield Inc. Quality Management System (QPM-7.6) which is ISO 9001:2015/AS9100 D certified and compliant with MIL-STD-45662A, ANSI/NCSS Z540.1-1994 (R2002) and ISO 10012.

This certificate applies only to item identified and shall not be reproduced without the permission from Barfield Inc.

Certified by: Freddy Dugué

Title/Stamp: Q.C. INSPECTOR **F.D. 01045**

Dated: 8/25/2025