

BARFIELD INC.

459-00024

MIAMI, FL • ATLANTA, GA • LONG BEACH, CA

FUEL QUANTITY T/S

P/N 101-00420

M/N 2548H

S/N 1124

MADE IN THE USA

MOD
A B C D E F G H J K
L M N O P Q R S T U

OPT
A B C D E F
G H J K L M

BARFIELD INC.
 4101 N.W. 29TH ST.
 Miami Fl - 33142
 (800) 321-1039



**CERTIFICATE OF CALIBRATION/
 FUNCTIONAL TEST**

W.O. No.: WO1233131

Customer Name PIUS PARTS

Date of Certification 9/11/2025 Certification Due Date: 9/11/2026
 Part Number: 101-00420 Received in Tolerance: NO
 Serial Number: 1124 Returned in Tolerance: YES
 Description: FUEL QUANTITY TEST SET Temperature: 74°
 Manufacturer: BARFIELD INC. Humidity: 43%

Calibration Procedure Used: 20-101-00420

Deviation: N/A

STANDARDS USED TO CERTIFY EQUIPMENT

Equipment	Serial Number.	Last Date Calibrated	Next Calibration Date
3729E	US36088159	6/25	6/26
2255E-1	B4786	6/25	6/26
2643EA	B8336	1/25	1/26
2715E	B8051	4/25	4/26
3064E	B8343	1/25	1/26
2762E-1	B8337	6/25	6/26

Comments UNIT FAILED TEST DUE TO OUT OF CALIBRATION.

Results: CALIBRATION FIXED FAULT. CALIBRATED AND TESTED UNIT W/I MFG. SPECS.

This is to certify that the above listed part number meets or exceeds all specifications as stated in the referenced procedure at the points tested (unless otherwise noted). Calibration Standards used in the calibration of this product are traceable to the National Institute of Standards and Technology (NIST). Unless otherwise stated the uncertainty of the measurement standards shall not exceed 25% of the acceptable tolerance. This calibration is in accordance with Barfield Inc. Quality Management System (QPM-7.6) which is ISO 9001:2015/AS9100 D certified and compliant with MIL-STD-45662A, ANSI/NCSL Z540.1-1994 (R2002) and ISO 10012.

This certificate applies only to item identified and shall not be reproduced without the permission from Barfield Inc.

Certified by: Freddy Duqué

Title/Stamp Q.C. INSPECTOR

F.D. 01045

Dated: 9/11/2025