

BARFIELD INC.  
4101 N.W. 29TH ST.  
Miami Fl - 33142  
(800) 321-1039



CERTIFICATE OF CALIBRATION/  
FUNCTIONAL TEST

W.O. No.: WO1256607

Customer Name PIUS PARTS

Date of Certification: 5/26/2026 Certification Due Date: 5/26/2027  
Part Number: 101-00540 Received in Tolerance: NO  
Serial Number: 104 Returned in Tolerance: YES  
Description: FUEL QTY. SYS. T/S Temperature: 74°  
Manufacturer: BARFIELD INC. Humidity: 43%

Calibration Procedure Used: 20-101-00540

Deviation: N/A

STANDARDS USED TO CERTIFY EQUIPMENT

Equipment	Serial Number.	Last Date Calibrated	Next Calibration Date
3729E	US36088159	06/25	06/26
2255E-1	B4786	06/25	06/26
2643EA	B8336	01/26	01/27
2762E-1	B8337	06/25	06/26
3064E-4	B0001	06/25	06/26
3064E-3	B0001	06/25	06/26
3064E-2	B8344	03/26	03/27
3064E	B8343	01/26	01/27
2715E	B8051	10/25	10/26
2255E-5	5785	06/25	06/26

Comments UNIT FAILS INS TEST DUE TO CRACKED INS TEST POINT SWITCH.

Results: REPLACED PART LISTED. CALIBRATE AND TESTED UNIT W/I MFG. SPECS.

This is to certify that the above listed part number meets or exceeds all specifications as stated in the referenced procedure at the points tested (unless otherwise noted). Calibration Standards used in the calibration of this product are traceable to the National Institute of Standards and Technology (NIST). Unless otherwise stated the uncertainty of the measurement standards shall not exceed 25% of the acceptable tolerance. This calibration is in accordance with Barfield Inc. Quality Management System (QPM-7.6) which is ISO 9001:2015/AS9100 D certified and compliant with MIL-STD-45662A, ANSI/NCSL Z540.1-1994 (R2002) and ISO 10012.

This certificate applies only to item identified and shall not be reproduced, without the permission from Barfield Inc.

Certified by: Claudia Galvez Title/Stamp: Q.C. INSPECTOR **C.G. 99910**